



Belegung

FS-OV-1

Dimensionen

25 m x 0,32 mm ID

Artikel Nr.

[232014](#)

Probe

Kakaobutter in n-Hexan

Substanzen: C50 / C52 / C54 / C56

Ofen


220-360°C mit 10°C/min

Trärgas

1,2 bar H<sub>2</sub> x Split 1:10

Detektor

FID 360°C, 1 x 512



1.) Triglyceride

